

Sample Name:GLOW70
Sample ID :
Time Processed: 1:47:09
Month-Day-Year Processed :07/01/2026

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm
<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	65
25.01	Pumps	B.Conc	65
27.00	Pumps	B.Conc	95
27.01	Pumps	B.Conc	5
35.00	Pumps	B.Conc	5
35.01	Controller	Stop	


<<Column Performance>>

<Detector A>

Column : Inertsil ODS-SP 4.6 x 250 mm
Equipment:ZJ19010014

<Chromatogram>

mV



<Peak Table>

Detector A Channel 1 220nm				
Peak#	Ret. Time	Area	Height	Area%
1	2.665	21150612	2214555	76.779
2	7.643	1521809	310433	5.524
3	10.096	4874935	557894	17.697
Total		27547356	3082882	100.000